



EDGEWOOD
CHEMICAL BIOLOGICAL CENTER
U.S. ARMY SOLDIER AND BIOLOGICAL CHEMICAL COMMAND

ECBC-TR-056

**CHARACTERIZATION OF MS2 BACTERIOPHAGE
BY INTEGRATED VIRUS DETECTION SYSTEM (IVDS)
PHYSICAL COUNTING METHODOLOGY**

Charles H. Wick

RESEARCH AND TECHNOLOGY DIRECTORATE


Patrick E. McCubbin

OPTIMETRICS, INC.
Forest Hill, MD 21050

August 1999

19990929 096

Approved for public release;
distribution is unlimited.



OptiMetrics, Inc.



Aberdeen Proving Ground, MD 21010-5424

DTIC QUALITY INSPECTED 4

Disclaimer

The findings in this report are not to be construed as an official Department of the Army position unless so designated by other authorizing documents.

| REPORT DOCUMENTATION PAGE | | | Form Approved OMB No. 0704-0188 | |
|--|---|--|--|--|
| Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503 | | | | |
| 1. AGENCY USE ONLY (Leave blank) | | 2. REPORT DATE 1999 August | | 3. REPORT TYPE AND DATES COVERED Final, 98 Mar - 98 Sep |
| 4. TITLE AND SUBTITLE Characterization of MS2 Bacteriophage by Integrated Virus Detection System (IVDS) Physical Counting Methodology | | | 5. FUNDING NUMBERS NONE | |
| 6. AUTHOR(S) Wick, Charles H. (ECBC), and McCubbin, Patrick E. (OPTIMETRICS) | | | | |
| 7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) DIR, ECBC,* ATTN: AMSSB-RRT-DS, APG, MD 21010-5424 OptiMetrics, Inc., 1 Newport Drive, Suite H, Forest Hill, MD 21050 | | | 8. PERFORMING ORGANIZATION REPORT NUMBER ECBC-TR-056 | |
| 9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) DIR, DARPA, Arlington, VA 22203 | | | 10. SPONSORING/ MONITORING AGENCY REPORT NUMBER | |
| 11. SUPPLEMENTARY NOTES *When this work was performed, ECBC was known as the U.S. Army Edgewood Research Development and Engineering Center (ERDEC). | | | | |
| 12A. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release; distribution is unlimited. | | | 12B. DISTRIBUTION CODE | |
| 13. ABSTRACT (Maximum 200 words) A new physically based methodology - The Integrated Virus Detection System (IVDS) - was used to characterize a high concentration, 10.2 mg protein/ml, sample preparation of MS2 Bacteriophage with a reported 10^{14} pfu/ml (DPM14) virus count in a common TNME buffer. Virus counts were made using the IVDS instrument following serial dilution. Results indicated virus counts of 1.5×10^5 for the neat sample (DPM14), followed by 6.5×10^4 viruses (DPM13), 1.2×10^4 viruses (DPM12), 9.3×10^2 viruses (DPM11), 88 viruses (DPM10), and 5 viruses (DPM9), respectively. Lower concentrations display a consistent multiplier and were consistent with target dilutions. Increases in virus concentration appear to decrease the multiplier. Variation is considered to be due to aggregation. Results demonstrate a consistent and simple-to-use methodology. Results further indicate that the IVDS instrument can be used for characterization of other virus preparations with equal ease and similar results. | | | | |
| 14. SUBJECT TERMS Virus Virus detection Separation Detection | | | 15. NUMBER OF PAGES 13 | |
| | | | 16. PRICE CODE | |
| 17. SECURITY CLASSIFICATION OF REPORT UNCLASSIFIED | 18. SECURITY CLASSIFICATION OF THIS PAGE UNCLASSIFIED | 19. SECURITY CLASSIFICATION OF ABSTRACT UNCLASSIFIED | 20. LIMITATION OF ABSTRACT UL | |

BLANK

PREFACE

The work described in this report was performed as part of a Defense Advanced Research Projects Agency (DARPA) project. This work was started in March 1998 and was completed in September 1998.

The use of either trade or manufacturers' names in this report does not constitute an official endorsement of any commercial products. This report may not be cited for purposes of advertisement.

This report has been approved for public release. Registered users should request additional copies from the Defense Technical Information Center; unregistered users should direct such requests to the National Technical Information Service.

Acknowledgments

Special thank you to Dr. Stephen S. Morse, DARPA (Arlington, VA), who funded the work, and to Linda Schaffer, Geo-Centers (Gunpowder Branch, APG), for her assistance in putting the manuscript together.

BLANK

CONTENTS

| | | |
|-----|--|----|
| 1. | Introduction | 7 |
| 2. | Laboratory Testing of MS2 Bacteriophage. | 8 |
| 2.1 | Results | 8 |
| 2.2 | Analysis | 12 |
| 3. | Conclusions | 13 |

FIGURES

| | | |
|----|---|----|
| 1. | MS2 1×10^{14} pfu/ml, DPG Lot #98110 . . . | 8 |
| 2. | Micrograph of MS2 bacteriophage . . . | 9 |
| 3. | MS2 1×10^{13} pfu/ml, DPG Lot #98110 . . . | 10 |
| 4. | MS2 1×10^{12} pfu/ml, DPG Lot #98110 . . . | 10 |
| 5. | MS2 1×10^{11} pfu/ml, DPG Lot #98110 . . . | 10 |
| 6. | MS2 1×10^{10} pfu/ml, DPG Lot #98110 . . . | 11 |
| 7. | MS2 1×10^9 pfu/ml, DPG Lot #98110 . . . | 11 |
| 8. | MS2 1×10^8 pfu/ml, DPG Lot #98110 . . . | 11 |

TABLES

| | | |
|----|--|----|
| 1. | Serial Dilution Samples of MS2 . . . | 9 |
| 2. | IVDS Physical Counts for MS2 Samples . . . | 12 |
| 3. | Numerical Analysis of MS2 Peak Count Information . . . | 13 |

Characterization of MS2 Bacteriophage by Integrated Virus Detection System (IVDS) Physical Counting Methodology

1. Introduction

The detection and analysis of viruses has been the goal of science for many years, following the first evidence that a new type of microorganism was responsible for diseases in both man and animals. These viruses were smaller than bacteria and thus presented the first challenge. Their small size made classification of these new microbes difficult and the field of virology was advanced by biochemical techniques rather than by direct examination. In more recent times and advancements in electronmicroscopy helped this problem and much information has been reported on the physical features of more than 21 virus families. These historic techniques are, however, time consuming, and require special knowledge, specialized chemicals or reagents and techniques to be successful. It was found possible based on the physical characteristics of viruses to count the individual viruses directly in a new and dramatic way using the Integrated Virus Detection System (IVDS)^{1,2} instrument, which uses easily obtained materials and simple to operate techniques.

This new instrument was used to analyze and characterize a sample of MS2 bacteriophage provided by the Life Sciences Division at Dugway Proving Ground (DPG). This sample was 2 ml of purified MS2 bacteriophage at a concentration 1×10^{14} plaque forming units (pfu)/ml or 10.2 mg protein/ml. This highly purified sample is from Lot #98110.

The MS2 sample was analyzed using the IVDS instrument or more directly the Gas-phase Electrophoretic Mobility Molecular Analyzer (GEMMA) detector which is one stage of the IVDS instrument. The GEMMA detector consists of an electrospray unit to inject samples into the detector, a Differential Mobility Analyzer and a Condensate Particle Counter. A complete description of the IVDS system, including the GEMMA detector, can be found in the report *Virus Detection: Limits and Strategies*.³

¹ Patent Pending on the IVDS technology.

² Wick, C.H., Anderson, D.M., and McCubbin, P.E., *Characterization of the Integrated Virus Detection System (IVDS) Using MS-2 Bacteriophage*, ECBC-TR-018, U.S. Army Edgewood Chemical Biological Center, Aberdeen Proving Ground, MD, May 1999.

³ Wick, C.H., Yeh, H.R., Carlon, H.R., and Anderson, D., *Virus Detection: Limits and Strategies*, ERDEC-TR-453, December 1997.

2. Laboratory Testing of MS2 Bacteriophage

2.1 Results

The high purity MS2 sample, with 1×10^{14} pfu/ml (hereafter described as DPM14) was analyzed using the GEMMA virus detector. The sample of DPM14 was placed neat into the GEMMA analyzer and the results are shown in Figure 1. The graph shows a very high virus count (over 150,000 counts) as well as other features. MS2 is nominally 24-26nm in size and this is illustrated in Figure 1. In fact, the sample as received was difficult to aspirate through the capillary delivery system in the GEMMA.




Figure 1 MS2 1×10^{14} pfu/ml, DPG Lot #98110

The size range of 24-26 nm is the expected size for a MS2 bacteriophage, as shown in Figure 2 in a micrograph by Dr. Hans Ackermann.⁴

⁴ Micrograph located at <http://life.anu.edu.au/viruses/welcome.htm>.




Figure 2 Micrograph of MS2 bacteriophage (bar represents 100 nm)

When the difficulty of sampling the neat MS2 sample became apparent, the sample DPM14 was then serially diluted to produce a number of lower concentration samples. That is, an aliquot of DPM14 was diluted 10 fold to produce a sample of MS2 at a concentration of 1×10^{13} pfu/ml. This sample was named DPM13. The dilutions were all made with a 0.02M solution of ammonium acetate (pH~10), which is required for the electrospray unit. The pH was adjusted to keep the virus from breaking down into its component subunits. Sample DPM13 was then diluted 10 fold, and likewise for the following dilutions. Table 1 lists the samples that were produced by serially dilution of the original sample.

Table 1 Serial Dilution Samples of MS2

| | |
|-------|---------------------------|
| DPM13 | 1×10^{13} pfu/ml |
| DPM12 | 1×10^{12} pfu/ml |
| DPM11 | 1×10^{11} pfu/ml |
| DPM10 | 1×10^{10} pfu/ml |
| DPM9 | 1×10^9 pfu/ml |
| DPM8 | 1×10^8 pfu/ml |

Figures 3-8 show the resultant GEMMA analysis of the serially diluted MS2 samples.




Figure 3 MS2 1×10^{13} pfu/ml, DPG Lot #98110




Figure 4 MS2 1×10^{12} pfu/ml, DPG Lot #98110




Figure 5 MS2 1×10^{11} pfu/ml, DPG Lot #98110




Figure 6 MS2 1×10^{10} pfu/ml, DPG Lot #98110




Figure 7 MS2 1×10^9 pfu/ml, DPG Lot #98110




Figure 8 MS2 1×10^8 pfu/ml, DPG Lot #98110

The counts for the serial dilutions were tabulated and are shown in Table 2.

Table 2 IVDS Physical Counts for MS2 Samples

| MS2 Sample | Counts in Size Range | | | | |
|------------|----------------------|-----------|-----------|-----------|-----------|
| | 25.946 nm | 25.029 nm | 24.144 nm | 23.291 nm | 22.468 nm |
| DPM8 | | | 1 | | |
| DPM9 | | 2 | 5 | 3 | |
| DPM10 | | 17 | 88 | 52 | |
| DPM11 | | 146 | 929 | 541 | 78 |
| DPM12 | 148 | 3613 | 12582 | 5174 | 255 |
| DPM13 | 15216 | 57624 | 65021 | 16893 | 1664 |
| DPM14 | 96995 | 157461 | 150886 | 65389 | 8347 |

2.2 Analysis

The GEMMA detector easily detects MS2 bacteriophage. The virus is consistently detected in the range of 22 to 26 nm. The GEMMA scans also show very low backgrounds away from the MS2 peaks. The action of serially diluting the MS2 did not affect the stability of the bacteriophage in solution. In fact, the addition of ammonium acetate buffer to produce dilutions reduced the background counts. The GEMMA scans of buffer solutions show very low counts, as ammonium acetate is nearly invisible to the detector.

The count rates for the various concentrations of MS2 were tabulated in Table 3. A comparison of the multiplication factor from sample to sample was also tabulated in the table. The lower concentrations display a fairly consistent multiplier and are consistent with the target dilutions. As the concentrations increase, the multiplier appears to decrease in magnitude. As was noted in Section 1.1, the as received sample, DPM14, was difficult to aspirate into the GEMMA detector. This sample is very concentrated and this appears to interfere with the analysis. The reduction in the multiplier may be caused by the agglomeration of particles as they flow through the Condensate Particle Counter (CPC) in the GEMMA unit. This agglomeration would lower the amount of particles counted and reduce the multiplier. It would appear that a count rate over 100,000 counts in a few adjacent channels, with a virus in this size range of 25 nm, is approaching an upper limit to concentrations that can be analyzed in the detector. This is easily remedied by simply diluting a sample to less than 100,000 counts in adjacent channels.

Table 3 Numerical Analysis of MS2 Peak Count Information

| MS2 Sample | Sum of size range | Multiplier from sample to sample |
|------------|-------------------|----------------------------------|
| DPM8 | 1 | - |
| DPM9 | 10 | 10.0 |
| DPM10 | 157 | 15.7 |
| DPM11 | 1694 | 10.8 |
| DPM12 | 21772 | 12.9 |
| DPM13 | 156418 | 7.2 |
| DPM14 | 479078 | 3.1 |

The actual sensitivity of the GEMMA detector was not in question in this study. The presented solution to the detector can be further concentrated to allow for the analysis of samples that appear to be too dilute. The sample DPM8 could be concentrated from one ml, the original volume, to 10 μ l. This would then present the GEMMA detector with a sample that would generate a graph with ~100 counts in a scan. The number of viruses that can be detected by the GEMMA is very low, on the order of 10 viruses, and therefore the ability to detect viruses is only a function of the presented solution concentration. A further example was a simple experiment where a few thousand viruses were measured into 500 ml of water. The water sample was concentrated through the Ultrafilter unit and nearly 800 viruses were counted by the GEMMA. The limiting factor for analysis is the ability to further concentrate a liquid solution while still being able to effectively handle the solution without losing it due the handling problems associated with tiny volumes.

3. Conclusions

The sample of MS2 bacteriophage received from the Life Sciences Division at Dugway Proving Ground was a very pure and concentrated sample. No other viruses were detected. The sample responded well to serial dilutions and was stable in the ammonium acetate buffer. This technique is a simple method to test the purity of any virus preparation since the IVDS instrument is not limited to any particular virus.